

TITLE

**METHOD AND DEVICE FOR PERFORMING
COOLING- OR CRYO-THERAPIES FOR, E.G.,
ANGIOPLASTY WITH REDUCED RESTENOSIS OR
PULMONARY VEIN CELL NECROSIS
TO INHIBIT ATRIAL FIBRILLATION**

CROSS-REFERENCE TO RELATED APPLICATIONS

10 This application is a divisional of U.S. Patent
Application Serial No. 09/787,599, filed 3/21/01, entitled
"Method And Device For Performing Cooling-Or Cryo-Therapies
For, E.G., Angioplasty With Reduced Restenosis Or Pulmonary
Vein Cell Necrosis To Inhibit Atrial Fibrillation" which is
15 a 371 National Phase Application claiming priority to PCT
Serial No. PCT/US01/06648, filed 3/1/01, entitled "Method
And Device For Performing Cooling-Or Cryo-Therapies For,
E.G., Angioplasty With Reduced Restenosis Or Pulmonary Vein
Cell Necrosis To Inhibit Atrial Fibrillation" which is a
20 continuation-in-part of U.S. Patent Application Serial No.
09/516,319, filed 3/1/00, entitled "Method and Device for
Performing Cooling-or Cryo-Therapies for, e.g., Angioplasty
with Reduced Restenosis or Pulmonary Vein Cell Necrosis to
Inhibit Atrial Fibrillation" which is a continuation-in-
25 part of U.S. Patent Application Serial No. 09/103,342,
filed 6/23/98, entitled "Selective Organ Cooling Catheter
And Method Of Using The Same" and of U.S. Patent
Application Serial No. 09/052,545, filed 3/31/98, entitled
"Circulating Fluid Hypothermia Method and Apparatus" and of
30 U.S. Patent Application Serial No. 09/215,038, filed
12/16/98, entitled "Inflatable Catheter for Selective Organ
Heating and Cooling and Method of Using the Same", all of
which are incorporated herein by reference.

BACKGROUND OF THE INVENTION

Balloon angioplasty, or the technology of reshaping of a blood vessel for the purpose of establishing vessel patency using a balloon tipped catheter, has been known since the late 1970's. The procedure involves the use of a balloon catheter that is guided by means of a guidewire through a guiding catheter to the target lesion or vessel blockage. The balloon typically is equipped with one or more marker bands that allow the interventionalist to visualize the position of the balloon in reference to the lesion with the aid of fluoroscopy. Once in place, i.e., centered with the lesion, the balloon is inflated with a biocompatible fluid, and pressurized to the appropriate pressure to allow the vessel to open.

Typical procedures are completed with balloon inflation pressures between 8 and 12 atmospheres. A percentage of lesions, typically heavily calcified lesions, require much higher balloon inflation pressures, e.g., upward of 20 atmospheres. At times, the balloon inflation procedure is repeated several times before the lesion or blockage will yield. The placement of stents after angioplasty has become popular as it reduces the rate of restenosis.

Restenosis refers to the renarrowing of the vascular lumen following vascular intervention such as a balloon angioplasty procedure or stent insertion. Restenosis is clinically defined as a greater than 50% loss of initial lumen diameter. The mechanism or root causes of restenosis are still not fully understood. The causes are multifactorial, and are partly the result of the injury caused by the balloon angioplasty procedure and stent placement. With the advent of stents, restenosis rates have

dropped from over 30% to 10-20%. Recently, the use and effectiveness of low-dose radiation administered intravascularly following angioplasty is being evaluated as a method to alter the DNA or RNA of an affected vessel's
5 cells in the hope of reducing cell proliferation.

Besides restenosis, another cardiological malady is atrial fibrillation. Atrial fibrillation refers to very rapid irregular contractions of the atria of the heart resulting in a lack of synchronization between the
10 heartbeat and the pulse. The irregular contractions are due to irregular electrical activity that originates in the area of the pulmonary veins. A proposed device, currently under development, for treating atrial fibrillation is a balloon filled with saline that can be ultrasonically
15 agitated and heated. This device is inserted in the femoral vein and snaked into the right atrium. The device is then poked through the interatrial septum and into the left atrium, where it is then angled into the volume adjoining the suspect pulmonary vein with the left atrium.

20 Research in atrial fibrillation indicates that substantially complete circumferential necrosis is required for a therapeutic benefit. The above technique is disadvantageous in that circumferential portions of the tissue, desired to be necrosed, are not in fact affected.
25 Other techniques, including RF ablation, are similarly inefficient. Moreover, these techniques leave the necrosed portions with jagged edges, i.e., there is poor demarcation between the healthy and the necrosed tissue. These edges can then cause electrical short circuits, and associated
30 electrical irregularities, due to the high electric fields associated with jagged edges of a conductive medium.

The above technique is also disadvantageous in that heating is employed. Heating is associated with several problems, including increased coagulum and thrombus formation, leading to emboli. Heating also stimulates stenosis of the vein. Finally, since tissues can only safely be heated to temperatures of less than or about 75°C - 85°C due to charring and tissue rupture secondary to steam formation. The thermal gradient thus induced is fairly minimal, leading to a limited heat transfer. Moreover, since heating causes tissues to become less adherent to the adjacent heat transfer element, the tissue contact with the heat transfer element is also reduced, further decreasing the heat transfer.

SUMMARY OF THE INVENTION

The present invention provides an enhanced method and device to inhibit or reduce the rate of restenosis following angioplasty or stent placement. The invention is similar to placing an ice pack on a sore or overstrained muscle for a period of time to minimize or inhibit the biochemical events responsible for an associated inflammatory response. An embodiment of the invention generally involves placing a balloon-tipped catheter in the area treated or opened through balloon angioplasty immediately following angioplasty. A so-called "cryoplasty" balloon, which can have a dual balloon structure, may be delivered through a guiding catheter and over a guidewire already in place from a balloon angioplasty. The dual balloon structure has benefits described below and also allows for a more robust design, providing significant safety advantages to the patient because two balloons must be broken if cooling fluid is to deleteriously infuse into the patient.

The dual balloon may be centered in the recently opened vessel with the aid of radio opaque marker bands, indicating the "working length" of the balloon. In choosing a working length, it is important to note that
5 typical lesions may have a size on the order of 2-3 cm. A biocompatible heat transfer fluid, which may contain contrast media, may be infused through the space between the dual balloons. While this fluid does not circulate in this embodiment, once it is chilled or even frozen by
10 thermal contact with a cooling fluid, it will stay sufficiently cold for therapeutic purposes. Subsequently, a biocompatible cooling fluid with a temperature between about, e.g., - 40°C and -60°C, may be injected into the interior of the inner balloon, and circulated through a
15 supply lumen and a return lumen. The fluid exits the supply lumen through a skive in the lumen, and returns to the refrigeration unit via another skive and the return lumen.

The biocompatible cooling fluid chills the
20 biocompatible heat transfer fluid between the dual balloons to a therapeutic temperature between about, e.g., 0°C and -50°C. The chilled heat transfer fluid between the dual balloons transfers thermal energy through the balloon wall and into the adjacent intimal vascular tissue for the
25 appropriate therapeutic length of time. Upon completion of the therapy, the circulation of the biocompatible cooling fluid is stopped, and the heat transfer fluid between the dual balloons withdrawn through the annular space. Both balloons may be collapsed by means of causing a soft vacuum
30 in the lumens. Once collapsed, the cryoplasty catheter may be withdrawn from the treated site and patient through the guiding catheter.

In more detail, in one aspect, the invention is directed to a device to treat tissue, including an outer tube, an an inner tube disposed at least partially within the outer tube, and a dual balloon including an inner
5 balloon and an outer balloon, the inner balloon coupled to the inner tube at a proximal and at a distal end, the outer balloon coupled to the inner tube at a distal end and to the outer tube at a proximal end. A first interior volume is defined between the outer balloon and the inner balloon
10 in fluid communication with an inlet in the volume between the outer tube and the inner tube.

Variations of the invention may include one or more of the following. The inner tube may further define a guidewire lumen, a supply lumen, and return lumen. The
15 supply lumen may define a hole or skive such that a fluid flowing in the supply lumen may be caused to flow into a volume defined by the inner balloon, and the return lumen may define a hole or skive such that a fluid flowing in a volume defined by the inner balloon may be caused to flow
20 into the return lumen. The guidewire lumen may extend from a proximal end of the inner tube to a distal end of the inner tube. The device may further comprise at least two radially extending tabs disposed around a circumference of the inner tube to substantially center the inner tube
25 within the dual balloon. The device may further comprise at least one marker band disposed on the inner tube to locate a working region of the device at a desired location. The device may further comprise a source of chilled fluid having a supply tube and a return tube, the
30 supply tube coupled in fluid communication to the supply lumen and the return tube coupled in fluid communication to the return lumen. A source of fluid may also be included,

the source of fluid coupled in fluid communication to a volume between the inner balloon and the outer balloon. The fluid may be a perfluorocarbon such as Galden fluid. The fluid may also include contrast media.

5 In another aspect, the invention is directed to a method of reducing restenosis after angioplasty in a blood vessel. The method includes inserting a catheter into a blood vessel, the catheter having a balloon. The balloon is then inflated with a perfluorocarbon such that an
10 exterior surface of the balloon is in contact with at least a partial inner perimeter of the blood vessel, the perfluorocarbon having a temperature in the range of about -10°C to -50°C .

Variations of the method may include one or more of
15 the following. The method may include disposing the catheter at a desired location using at least one radio opaque marker band. The method may include flowing the perfluorocarbon into the balloon using a supply lumen and exhausting the perfluorocarbon from the balloon using a
20 return lumen. The balloon may be a dual balloon, and the method may further include providing a heat transfer fluid in the volume between the dual balloons. The heat transfer fluid may include a contrast media fluid. The method may include disposing the catheter such that at least a portion
25 of the balloon is in a coronary artery or in a carotid artery.

In yet another aspect, the invention is directed to a method of reducing atrial fibrillation. The method includes inserting a catheter at least partially into the
30 heart, the catheter having a balloon, a portion of the balloon located in the left atrium and a portion of the balloon located in a pulmonary vein. The balloon is then

inflated with a perfluorocarbon such that an exterior surface of the balloon is in contact with at least a partial circumference of the portion of the pulmonary vein adjacent the left atrium, the perfluorocarbon having a
5 temperature in the range of about -10°C to -50°C .

Variations of the method may include one or more of the following. The balloon may have a working region having a length of between about 5 mm and 10 mm. The method may further include inserting a wire having a needle point
10 from the femoral vein into the right atrium and forming a hole using the needle point in the interatrial septum between the right atrium and the left atrium. A guide catheter may then be inserted into the right atrium. A guide wire may further be inserted through the guide
15 catheter into the right atrium and further into a pulmonary vein. The catheter may then be disposed over the guidewire into a volume defined by the joint of the right atrium and the pulmonary vein.

Advantages of the invention may include one or more of
20 the following. The invention inhibits or reduces the rate of restenosis following a balloon angioplasty or any other type of vascular intervention. At least the following portions of the vascular anatomy can benefit from such a procedure: the abdominal aorta (following a stent or graft
25 placement), the coronary arteries (following PTCA or rotational artherectomy), the carotid arteries (following an angioplasty or stent placement), as well as the larger peripheral arteries.

When the invention is used to treat atrial
30 fibrillation, the following advantages inure. The cooled tissue is adherent to the heat transfer element, increasing the heat transfer effected. Since very cold temperatures

may be employed, the temperature gradient can be quite large, increasing the heat transfer rate.

In both embodiments, heat transfer does not occur primarily or at all by vaporization of a liquid, thus
5 eliminating a potential cause of bubbles in the body. Nor does cooling occur primarily or at all by a pressure change across a restriction or orifice, this simplifying the structure of the device. Thrombus formation and charring, associated with prior techniques, are minimized or
10 eliminated.

Additional advantages will be apparent from the description that follows, including the drawings and claims.

15 BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1A shows a side schematic view of a catheter according to a first embodiment of the invention.

FIG. 1B shows a cross-sectional view of the catheter of FIG. 1A, as indicated by lines 1B-1B in FIG. 1A.

20 FIG. 1C shows an alternate cross-sectional view of the catheter of FIG. 1A, as indicated by lines 1B-1B in FIG. 1A.

FIG. 2A shows a side schematic view of a catheter according to a second embodiment of the invention.

25 FIG. 2B shows a cross-sectional view of the catheter of FIG. 2A, as indicated by lines 2B-2B in FIG. 2A.

FIG. 3 shows a schematic view of a catheter in use according to a third embodiment of the invention.

FIG. 4 shows a cross-sectional view of the catheter of
30 FIG. 3.

FIG. 5 shows an alternative cross-sectional view of the catheter of FIG. 3.

FIG. 6 shows an alternative cross-sectional view of the catheter of FIG. 3.

FIG. 7 shows a schematic view of the warm balloon of the catheter of FIG. 3.

5

DETAILED DESCRIPTION

Referring to FIG. 1A, a catheter 100 is shown according to a first embodiment of the invention. The catheter 100 has a proximal end 130 and a distal end 114. Of course, this figure is not necessarily to scale and in general use the proximal end 130 is far upstream of the features shown in FIG. 1A.

The catheter 100 may be used within a guide catheter 102, and generally includes an outer tube 103, a dual balloon 134, and an inner tube 122. These parts will be discussed in turn.

The guide catheter 102 provides a tool to dispose the catheter 100 adjacent the desired location for, e.g., angioplasty or reduction of atrial fibrillation. Typical guide catheter diameters may be about 6 french to 9 french, and the same may be made of polyether blockamide, polyamides, polyurethanes, and other similar materials. The distal end of the guide catheter is generally adjacent the proximal end of the dual balloon 134, and further is generally adjacent the distal end of the outer tube 103.

The ability to place the guide catheter is a significant factor in the size of the device. For example, to perform angioplasty in the carotid arteries, which have an inner diameter of about 4 to 6 mm, a suitably sized guide catheter must be used. This restricts the size of the catheter 100 that may be disposed within the guide

catheter. A typical diameter of the catheter 100 may then be about 7 french or less or about 65 to 91 mils. In a second embodiment described below, a catheter for use in the coronary arteries is described. Of course, which
5 catheter is used in which artery is a matter to be determined by the physician, taking into account such factors as the size of the individual patient's affected arteries, etc.

The outer tube 103 houses the catheter 100 while the
10 latter traverses the length of the guide catheter 102. The outer tube 103 may have a diameter of about 4 french to 7 french, and the same may be made of polyether blockamide, poly-butylene terephthalate, polyurethane, polyamide, polyacetal polysulfone, polyethylene, ethylene
15 tetrafluoroethylene, and other similar materials.

The distal end of the outer tube 103 adjoins the proximal end of the dual balloon 134. The outer tube 103 provides a convenient location for mounting a proximal end of an outer balloon 104 within the dual balloon 134, and
20 further may provide an inlet 128 for providing a fluid such as a liquid to a first interior volume 106 between the dual balloons. In some cases, an inlet 128 per se may not be necessary: the fluid, which may also be a sub-atmospheric level of gas or air, may be provided during manufacture in
25 the first interior volume 106. In this case, the proximal and distal ends of the first interior volume may be sealed during manufacture. The inlet 128 may be at least partially defined by the annular volume between the interior of the outer tube 103 and the exterior of the
30 inner tube 122.

The dual balloon 134 includes an outer balloon 104 and an inner balloon 108. Between the two is the first

interior volume 106. The outer balloon 104 may be inflated by inflating the interior volume 106. The inner balloon 108 has a second interior volume 110 associated with the same. The inner balloon 108 may be inflated by inflating
5 the second interior volume 110.

To avoid the occurrence of bubbles in the bloodstream, both the inner balloon 108 and the outer balloon 104 may be inflated using biocompatible liquids, such as Galden® fluid, perfluorocarbon-based liquids, or various contrast
10 agents. There is no need that the fluid inflating one of the interior volumes be the same fluid as that inflating the other. Additional details on these fluids are described below.

In the case of the first interior volume 106, this
15 fluid may be, e.g., stationary or static: in other words, it need not be circulated. In the case of the second interior volume 110, this fluid would in general be circulated by an external chiller (not shown). The chiller may be, e.g., a gear pump, peristaltic pump, etc. It may
20 be preferable to use a gear pump over a peristaltic pump as the attainable pressure of the former is generally greater than that of the latter. Moreover, gear pumps have the advantageous property of being linear, i.e., their output varies in direction proportion with their revolutions per
25 minute. Two types of gear pumps which may be employed include radial spur gear pumps and helical tooth gear pumps. Of these, the helical tooth gear pump may be more preferable as the same has been associated with higher pressures and a more constant output. The ability to
30 achieve high pressures may be important as the cooling fluid is required to pass through a fairly narrow, e.g., five to seven french, catheter at a certain rate. For the

same reason, the viscosity of the fluid, at the low temperatures, should be appropriately low. In this way, e.g., the flow may be increased. For example, an appropriate type of fluid may be Galden® fluid, and in particular Galden® fluid item number "HT-55", available from Ausimont Inc. of Thorofare, NJ. At -55°C, this fluid has a viscosity of 2.1 centiStokes. At -70°C, this fluid has a viscosity of 3.8 centiStokes. It is believed that fluids with such viscosities at these temperatures would be appropriate for use.

The so-called "cones" of the balloons 108 and 104, indicated generally by reference numeral 132, may be made somewhat thicker than the remainder of the balloon sections. In this way, the heat transfer efficiency in these sections is significantly less than over the remainder of the balloon sections, this "remainder" effectively defining a "working region" of the balloon. In this way, the cooling or "cryoplasty" may be efficiently localized to the affected area rather than spread over the length of the balloon.

The inner tube 122 is disposed within the interior of the dual balloon 134 and within the interior of the guide catheter 102. The inner tube 122 includes a supply lumen 120, a return lumen 118, and a guidewire lumen 116. The guidewire lumen 116 may have sizes of, e.g., 17 or 21 mils inner diameter, in order to accommodate current standard sized guidewires, such as those having an outer diameter of 14 mils. This structure may be preferable, as the pressure drop encountered may be substantially less. In use, the supply lumen 120 may be used to supply a circulating liquid to the second interior volume 110. The return lumen 118 may be used to exhaust the circulating liquid from the

second interior volume to the external chiller. As may be seen from FIG. 1A, both lumens 118 and 120 may terminate prior to the distal end 114 of the catheter 100. The lumen arrangement may be seen more clearly in Fig. 1B. Fig. 1C shows an alternate such arrangement, and one that may provide an even better design for minimal pressure drop. In this design, lumens 118' and 120' are asymmetric about guidewire lumen 116'.

A set of radio opaque marker bands 112 may be disposed on the inner tube 122 at locations substantially adjacent the cones 132 to define a central portion of the "working region" of the balloons 104 and 108. This working region is where the "cryoplasty" procedures described below may substantially occur.

As noted above, the proximal portion of the outer balloon 104 is mounted on the outer tube 103 at its distal end. The distal end of the outer balloon 104 is secured to the distal end of the catheter 100 and along the inner tube 122. In contrast, both the proximal and distal ends of the inner balloon 108 may be secured to the inner tube 122 to create a sealed second interior volume 110.

At least two skives 124 and 126 may be defined by the inner tube 122 and employed to allow the working fluid to exit into the second interior volume 110 and to exhaust the same from the second interior volume 110. As shown in the figure, the skive 124 is in fluid communication with the lumen 120 and the skive 126 is in fluid communication with the lumen 118. Here, "fluid communication" refers to a relationship between two vessels where a fluid pressure may cause a net amount of fluid to flow from one vessel to the other.

The skives may be formed by known techniques. A suitable size for the skives may be from about 50 mils to 125 mils.

A plurality of tabs 119 may be employed to roughly or substantially center the inner tube 122 within the catheter 100. These tabs may have the shape shown, the shape of rectangular or triangular solids, or other such shapes so long as the flow of working fluid is not unduly impeded. In this specification, the phrase "the flow of working fluid is not unduly impeded" is essentially equated to the phrase "substantially center". The tabs 119 may be made of polyether blockamide, poly-butylene terephthalate, polyurethane, polyamide, polyacetal polysulfone, polyethylene, ethylene tetrafluoroethylene, and other similar materials, and may have general dimensions of from about 3 mils to 10 mils in height, and by about 10 mils to 20 mils in width.

In a method of use, the guide catheter 102 may be inserted into an affected artery or vein such that the distal tip of the guide catheter is just proximal to an affected area such as a calcified area or lesion. Of course, it is noted that typical lesions do not occur in the venous system, but only in the arterial.

This step provides a coarse estimate of proper positioning, and may include the use of fluoroscopy. The guide catheter may be placed using a guide wire (not shown). Both the guide catheter and guide wire may already be in place as it may be presumed a balloon angioplasty or stent placement has previously been performed.

The catheter 100 may then be inserted over the guide wire via the lumen 116 and through the guide catheter 102. In general, both a guide wire and a guide catheter are not

strictly necessary - one or the other may often suffice.

During insertion, the dual balloon 134 may be uninflated to maintain a minimum profile. In fact, a slight vacuum may be drawn to further decrease the size of the dual balloon 134 so long as the structural integrity of the dual balloon 134 is not thereby compromised.

When the catheter 100 is distal of the distal tip of the guide catheter 102, a fine positioning step may occur by way of the radio opaque marker bands 112. Using fluoroscopy, the location of the radio opaque marker bands 112 can be identified in relation to the location of the lesion. In particular, the catheter may be advantageously placed at the location of the lesion and further such that the lesion is between the two marker bands. In this way, the working region of the balloon 134 will substantially overlap the affected area, i.e., the area of the lesion.

Once placed, a biocompatible heat transfer fluid, which may also contain contrast media, may be infused into the first interior volume 106 through the inlet 128. While the use of contrast media is not required, its use may allow early detection of a break in the balloon 104 because the contrast media may be seen via fluoroscopy to flow throughout the patient's vasculature. Subsequently a biocompatible cooling fluid may be circulated through the supply lumen 120 and the return lumen 118. Before or during the procedure, the temperature of the biocompatible cooling fluid may be lowered to a therapeutic temperature, e.g., between -40°C and -60°C, although the exact temperature required depends on the nature of the affected area. The fluid exits the supply lumen 120 through the skive 124 and returns to the chiller through the skive 126 and via the return lumen 118. It is understood that the

respective skive functions may also be reversed without departing from the scope of the invention.

The biocompatible cooling fluid in the second interior volume 110 chills the biocompatible heat transfer fluid within the first interior volume 106 to a therapeutic temperature of, e.g., between about -25°C and -50°C . The chilled heat transfer fluid transfers thermal energy through the wall of the balloon 104 and into the adjacent intimal vascular tissue for an appropriate therapeutic length of time. This time may be, e.g., about $\frac{1}{2}$ to 4 minutes.

Upon completion of the therapy, the circulation of the biocompatible cooling fluid may cease. The heat transfer fluid within the first interior volume 106 may be withdrawn through the inlet 128. The balloons 104 and 108 may be collapsed by pulling a soft vacuum through any or all of the lumens 124, 126, and 128. Following collapse, the catheter 100 may be withdrawn from the treatment site and from the patient through the guide catheter 102.

To inhibit restenosis, the following therapeutic guidelines may be suggested:

	Minimum	Average	Maximum
Temperature of heat transfer fluid	-20°C	-55°C	-110°C
Temperature achieved at intimal wall	0°C to -10°C	-20°C to -30°C	-50°C to -100°C
Depth of penetration of intima/media	10ths of mm	1 mm	3 mm
Length of time fluid is circulating	30 seconds	1-2 min	4-5 min

5 Substantially the same catheter may be used to treat atrial fibrillation. In this method, the catheter is inflated as above once it is in location. The location chosen for treatment of atrial fibrillation is such that the working region spans a portion of the left atrium and a
10 portion of the affected pulmonary vein. Thus, in this embodiment, the working region of the catheter may have a length of about 5 mm to 30 mm. The affected pulmonary vein, of the four possible pulmonary veins, which enter the left atrium, may be determined by electrophysiology studies.

15 To maneuver the catheter into this location, a catheter with a needle point may first be inserted at the

femoral vein and routed up to the right atrium. The needle of the catheter may then be poked through the interatrial septum and into the left atrium. The catheter may then be removed if desired and a guide catheter disposed in the same location. A guide wire may be used through the guide catheter and may be maneuvered at least partially into the pulmonary vein. Finally, a catheter such as the catheter 100 may be placed in the volume defining the intersection of the pulmonary vein and the left atrium.

A method of use similar to that disclosed above is then employed to cool at least a portion of, and preferably all of, the circumferential tissue. The coldness of the balloon assists in the adherence of the circumferential tissue to the balloon, this feature serving to increase the overall heat transfer rate.

The catheter 100 above may be particularly useful for procedures in the carotid arteries by virtue of its size. For use in the coronary arteries, which are typically much smaller than the carotid artery, an even smaller catheter may be desired. For example, one with an outer diameter less than 5 french may be desired.

Referring to FIG. 2A, a catheter 200 is shown according to a second embodiment of the invention. This embodiment may be particularly useful for use in the coronary arteries because the dimensions of the catheter 200 may be considerably smaller than the dimensions of the catheter 100. However, in several ways the catheter 200 is similar to the above-described catheter 100. In particular, the catheter 200 has a proximal end 230 and a distal end 214 and may be used within a guide catheter 202. The catheter 200 includes an outer tube 203, a dual balloon 234, and an inner tube 222.

The ability to place the guide catheter is a significant factor in the size of the device. For example, to perform angioplasty in the coronary arteries, which have an inner diameter of about 1 ½ to 4 ½ mm, a suitably sized guide catheter may be used. This then restricts the size of the catheter 200 which may be disposed within the guide catheter. A typical diameter of the catheter 200 may then be about 3 french or less or about 35-39 mils. The same may be placed in the femoral artery in order to be able to track to the coronary arteries in a known manner.

Analogous to these features in the catheter 100, the outer tube 203 houses the catheter 200 and may have an outside diameter of about 5 french to 7 french, and the same may be made of similar materials. The distal end of the outer tube 203 adjoins the proximal end of the dual balloon 234. The outer tube 203 provides a mounting location for an outer balloon 204, and further provides an inlet 228 for providing a fluid such as a liquid to a first interior volume 206 between the dual balloons. As noted in connection with catheter 100, an inlet 228 per se may not be necessary: the fluid, which may also be a sub-atmospheric level of air, may be provided in the first interior volume 206. Also as above, the proximal and distal ends of the volume may be sealed during manufacture. The inlet 228 may be at least partially defined by the annular volume between the interior of the outer tube 203 and the exterior of the inner tube 222.

The dual balloon 234 includes an outer balloon 204 and an inner balloon 208. These balloons are basically similar to balloons 104 and 108 described above, but may be made even smaller for use in the smaller coronary arteries.

The same types of fluids may be used as in the catheter 100.

The inner tube 222 is disposed within the interior of the dual balloon 234 and within the interior of the guide
5 catheter 202. The inner tube 222 includes a supply lumen 220 and a return lumen 218.

A set of radio opaque marker bands 212 may be disposed on the inner tube 222 for the same reasons disclosed above in connection with the marker bands 112.

10 As noted above, the proximal portion of the outer balloon 204 is mounted on the outer tube 203 at its distal end. The distal end of the outer balloon 204 is secured to the distal end of the catheter 200 and along the inner tube 222. In contrast, both the proximal and distal ends of the
15 inner balloon 208 may be secured to the inner tube 222 to create a sealed second interior volume 210.

At least two skives 224 and 226 may be defined by the inner tube 222 and employed to allow the working fluid to exit into the second interior volume 210 and to exhaust the
20 same from the second interior volume 210.

A plurality of tabs 219 may be employed to roughly or substantially center the inner tube 222 within the catheter 200 as in catheter 100. These tabs may have the same general geometry and design as tabs 119. Of course, they
25 may also be appropriately smaller to accommodate the smaller dimensions of this coronary artery design.

The tabs 119 and 219 are particularly important in the catheters 100 and 200, as contact by the inner tube of the outer tube may also be associated with an undesired
30 conductive heat transfer prior to the working fluid reaching the working region, thereby deleteriously

increasing the temperature of the working fluid at the working region.

The method of use of the catheter 200 is generally the same as for the catheter 100. Known techniques may be employed to place the catheter 200 into an affected coronary artery. For the catheter 200, an external guidewire may be used with appropriate attachments to the catheter.

Referring to Fig. 3, an alternative embodiment of a catheter 300 which may be employed in PV ablation is detailed. In this figure, a dual balloon system 301 is shown; however, the balloons are not one within the other as in Fig. 1. In this embodiment, a warm balloon 302 is distal of a cold balloon 304. Warm balloon 302 may be used to anchor the system 301 against movements, which may be particularly useful within a beating heart. Cold balloon 304 may then be employed to cryo-ablate a circumferential lesion at the point where a pulmonary vein 306 enters the left atrium 308.

Within the cold balloon 304, a working fluid may be introduced via an outlet port 308 and may be retrieved via an inlet port 310. Ports 308 and 310 may be skived in known fashion into the catheter shaft lumens whose design is exemplified below.

As noted above, the warm balloon 302 serves to anchor the system 301 in the pulmonary vein and left atrium. The warm balloon 302 also serves to stop blood, which is traveling in the direction indicated by arrow 312, from freezing upon contact with the cold balloon 304. In this way, the warm balloon 302 acts as an insulator to cold balloon 304.

As the warm balloon 302 does not require convective heat transfer via a circulating working fluid, it may be served by only one skived port, or by two ports, such as an inlet port 314 and an outlet port 316, as shown in Fig. 3.

5 In some embodiments, a separate lumen or lumens may be used to fill the warm balloon. In an alternative embodiment, a valve mechanism may be used to fill the warm balloon using fluid from the cold balloon. In the case where only one port is used to fill the warm balloon, draining the same
10 requires a slight vacuum or negative pressure to be placed on the lumen servicing the inlet/outlet port. A benefit to the two lumen design is that the warm balloon may be inflated and deflated in a more expeditious manner.

Typical pressures within the warm balloon may be about
15 1-2 atm (10-30 psi), and thus maintains a fairly low pressure. An appropriate fluid will be biocompatible, and may be Galden fluid, D5W, and so on. Typical pressures within the cold balloon may be about 5-7 atm, for example about 6 atm (e.g., at about 100 psi), and thus maintains a
20 higher pressure. An appropriate fluid may be Galden fluid, e.g., HT-55, D5W, and so on. The volume of fluid required to fill the cold balloon may vary, but may be about 4-8 cc. The cold balloon may be about 2 to 2 ½ cm long, and have a diameter of 1 to 2 ½ cm.

25 In some embodiments, the warm balloon may be glued or otherwise attached to the cold balloon. In the case where only one port is used to fill the warm balloon, draining both balloons may simply entail closing either the return lumen or the supply lumen, and drawing a vacuum on the
30 other. In this way, both the cold and warm balloons may be evacuated. In any case, a standard medical "indeflator"

may be used to pressurize and de-pressurize the various lumens and balloons.

Fig. 4 shows an embodiment of the arrangement of lumens within the catheter. In particular, supply and return lumens for the cold balloon 304 are shown by lumens 318 and 320, respectively. Supply and return lumens for the warm balloon 302 are shown by lumens 322 and 324, respectively, although as noted only one may be used as required by the dictates of the user. A guidewire lumen 326 is also shown. An alternative arrangement is shown in Fig. 5, where the corresponding lumens are shown by primes.

In the above lumen designs, the exterior blood is exposed to the cold supply flow. Referring to Fig. 6, an alternative lumen design is shown in which the cold fluid supply lumen 328 is exposed to only the cold fluid return lumen 330. An insulation space 332 may also be employed. In this way, the heat flux from the exterior flow is minimized and the cold fluid may reach the cold balloon at a lower temperature. One drawback to such a system is that the operational pressure may be higher.

Referring back to Fig. 4, the overall catheter outer diameter may be about .130", e.g. about 10 french, including an insulation sleeve and guide discussed below. The catheter shaft 303 itself may be about .110" and may be made of, e.g., polyethylene (PE), and preferably a combination of a low density PE and a high density PE.

The inlet and outlet ports or inlet/outlet port of the warm balloon may be skived from the lumens 322 and 324. Referring to Fig. 7, the warm balloon 302 itself may be made of a sleeve 332 of silicone tubing of, e.g., 35 durometer on the "D" scale, and held in place by two pieces of PET heat shrink tubing 334. Alternative methods of

securing the warm balloon during inflation may include metal bands or an adhesive.

Referring back to Fig. 3, marker bands 336 may be employed within either or both of the cold balloon and warm
5 balloon to assist the physician in disposing the same in an appropriate location. The marker bands typically denote the working areas of the balloons, and may be made of Pt, Iridium, Au, etc.

In the ablation procedure, the working cold fluid may
10 exit the circulation system or chiller at, e.g., about -85°C . The circulation system or chiller may be, e.g., a two-stage heat exchanger. The fluid may then enter the catheter at about -70°C to about -75°C , and may strike the balloon at about -55°C to about -65°C . The overall
15 procedure may take less than a minute to circumferentially ablate the desired tissue up to several minutes. Of course, these numbers are only exemplary and the same depend on the design of the system and fluids used.

Mapping electrodes 338 may be employed at the distal
20 end of the warm balloon. These mapping electrodes may each have a wire attached, the wires extending down, e.g., the supply and return lumens for the warm fluid or the cold fluid. The mapping electrodes 338 may be used to detect stray electrical fields to determine where ablation may be
25 needed and/or whether the ablation procedure was successful. The mapping electrodes may typically be about 2-3 mm apart from each other.

Construction of the warm balloon typically involves adhering the same to the shaft 303 and skiving the inlet
30 and outlet ports. In some instances, it may be desired to place a silicone sleeve 340 on the proximal and/or distal ends of the warm and/or cold balloons. The silicone sleeve

340 may then serve to further insulate the non-working sections of the balloons from blood that would otherwise potentially freeze during a procedure. The silicone sleeve would typically be attached only at a portion of its length, such as that indicated by circle 342, so that the same may slide along the balloon as the balloon is inflated. In addition to insulation effects, the silicone sleeve also serves to assist in collapsing the balloon during deflation.

The entire catheter shaft 303 may be surrounded by an insulation catheter sleeve 344 (see Fig. 4). Sleeve 344 may have a thickness of, e.g., 0.01 inches, and may be made of a foamed extrusion, e.g., that with voids of air disposed within. The voids further assist the insulating effect since their heat transfer is extremely low. A void to polymer ratio of, e.g., 20% to 30% may be particularly appropriate. Such foamed extrusions are available from, e.g., Applied Medical Resources in Laguna Niguel, CA, or Extrusioneering, Inc., in Temecula, CA.

To prevent damage to tissue other than where the ablation is to occur, such as at the insertion site near the femoral vein and around the puncture point through the atrial septum, an insulation sleeve may be used as noted above.

The invention has been described above with respect to particular embodiments. It will be clear to one of skill in the art that numerous variations may be made from the above embodiments with departing from the spirit and scope of the invention. For example, the invention may be combined with stent therapies or other such procedures. The dual balloon disclosed may be used after angioplasty or may be an angioplasty balloon itself. Furthermore, while

the invention has occasionally been termed herein a "cryoplasty catheter", such a term is for identification purposes only and should not be viewed as limiting of the invention. Fluids that may be used as heat transfer fluids include perfluorocarbon-based liquids, i.e., halogenated hydrocarbons with an ether bond, such as FC 72. Other materials that may be used include CFCs, Freon®, or chemicals that when placed together cause an endothermic reaction. Preferably, low viscosity materials are used as these result generally in a lessened pressure drop. The balloons may be made, e.g., of Pebax, PET/PEN, PE, PA 11/12, PU, or other such materials. Either or both of the dual balloons may be doped to improve their thermal conductivities. The shaft of inner tube 122 may be made of Pebax, PBT, PI/PEI, PU, PA 11/12, SI, or other such materials. The precise shapes and dimensions of the inner and outer lumens, while indicated in, e.g., Figs. 1B, 1C, and 2B, may vary. The lumen design shown in Figs. 1B-1C may be employed in the catheter of Fig. 2A and vice-versa. Embodiments of the invention may be employed in the field of cold mapping, where a circle of tissue is cooled to see if the affected part has been reached. If the affected tissue is that which is being cooled, a more vigorous cooling may be instituted. Other variations will be clear to one of skill in the art, thus the invention is limited only by the claims appended hereto.

WHAT IS CLAIMED IS: